



Goddard Laser for Absolute Measurement of Radiance for Instrument Calibration in the Ultraviolet to Short Wave Infrared

Brendan McAndrew[a], Joel McCorkel[a], Timothy Shuman[b], Barbara Zukowski[c], Aboubakar Traore[d], Michael Rodriguez[e], Steven Brown[f], and John Woodward[f]

[a]NASA Goddard Space Flight Center, Greenbelt, MD, USA 20771

[b]Fibertek, Herndon, VA, USA 20171

[c]Ball Aerospace, Lanham, MD, USA 20706

[d]NASA Langley Research Center, Hampton, VA, USA 23666

[e]Sigma Space Corporation, Lanham, MD, USA 20706



[f]National Institute of Standards and Technology, Gaithersburg, MD, USA 20899





Introduction

- Purpose
- Absolute radiometric scale
- Calibration scheme
- Narrow linewidth tunable sources



Light source



radiometry



prelaunch calibration for instruments



Purpose

- Narrow linewidth, high spectral bandwidth sources provide higher signal and dynamic range, improved wavelength and radiance accuracy over broadband lamp based techniques
- More straightforward measurement and data interpretation – flat field, full signal level
- Enables more science, reduced mission lifetimes & cost





Absolute radiometric scale


Radiance: power per unit area per unit solid angle $L = \frac{P}{A \cdot \Omega}$

Spectral radiance: radiance per unit wavelength $L_\lambda = L / \Delta\lambda$


Greatest uncertainty is in optical power P

Area and solid angle are both traceable to meters

Optical power measured with electrical substitution radiometer and traceable to electrical units of measure




Narrow linewidth source eliminates error due to convolution of source spectrum with radiometer responsivity




Integrating sphere with transfer radiometers

Traceability path

NIST Facility



POWR
Primary Optical Watt
Radiometer



Stabilized laser source is used to transfer radiometric scale from POWR to portable transfer radiometer via another standard radiometer



LTD-11 #107
transfer radiometer

Sensor vendor facility



LTD-11 #107
transfer radiometer



Sphere Monitor



Satellite/airborne sensor




Tunable sources

Desired properties:

- **Radiometrically stable**
- Wavelength range covering the solar spectrum, 300 to 2500 nm
- Signal level comparable to maximum reflected solar radiance (snow and cloud cover at high sun angles)
- Linewidth \ll instrument under test
- Portability, minimal setup and facility infrastructure requirements
- Time efficiency: automated wavelength tuning, synchronized tuning with shutter cycling, instrument data collects
- Reliability: critical path operations

Solar spectrum



Lamp/monochromator
Titanium sapphire
Dye
Tunable diode
Optical parametric oscillator



Sources setup for JPSS-2

Dye

570 to 680 nm


Ti:sapphire

680 to 1000 nm

OPO

Second harmonic: 360 to 570 nm
Signal and idler: 720 to 2000 nm

Joint Polar Satellite System 2
Visible and Infrared Imaging
Radiometer Suite calibration




Combination of tunable
sources covering 360 to
2000 nm




Custom LBO OPO


NIR-OPO 680-1100 nm + 1200-2200 nm



SWIR-OPO 1080-1400 nm + 540-700 nm



532 nm pumped LBO OPO temperature tuning (e \rightarrow o+o)






Automation

Real time display + recording of wavelength, radiance, shutter state, and OPO parameters

Light and dark dwell time


Scan wavelength interval

Automated tuning via parameter look up table





Example calibration data




VIIRS band M2 showing electronic crosstalk with other bands



Quasi-CW & CW sources

Integrating sphere provides temporal as well as spatial averaging with characteristic time on the order of nanoseconds



Comparison of 80 MHz prf OPO and continuous wave Ti:sapphire laser used as sources for VIIRS band M7 calibration



Future work

- Deeper UV coverage
- Improved operation near degeneracy and water vapor absorption lines
- More repeatable wavelength steps, order of 0.1 nm
- Long term stability and decreased operator intervention during scans
- More power



Summary, acknowledgements

Broad band optical parametric oscillator developed covering 360 to 2000 nm
Automated scanning
Demonstrated portability without realignment
Ability to support critical path instrument calibration

Technical Support

Based on design work and built with assistance from the late Keith Lykke, NIST

Support

GOES-R

NPP

SAGE III – ISS

Landsat

PACE Ocean Color Instrument

CLARREO Pathfinder

Joint Polar Satellite System



PATHFINDER

